



1. Copy the .up file into U disk (FAT32 format).
2. Power on. Connect DSO3000 to PC by USB cable. Insert U disk into “USB HOST”.




3. Click “IO->Connection Expert” to open Keysight Connection Expert.



4. Click “Rescan”.
5. Connect device successfully. Click “Send Commands To This Instrument”.



6. Enter the next interface. Input "UP". Click "Send Command" to start update.



7. The four indicator lights will flicker. Wait for program.
8. When complete, there will be a buzzer sound and the device will restart. Observe the interface of Keysight Connection Expert, the device will be reconnected.
9. That's all.